

Certificate of Conformity

Number: C223-0010/16 **Project file:** C20160009

Product: Dry Salt Aerosol Generator for Inhalation

Type reference: PRIZSALT+

Trademark:



Applicant: PRIZMA d.o.o.
Kumanovska 8, RS-34000 Kragujevac, Serbia

Manufacturer: PRIZMA d.o.o.
Kumanovska 8, RS-34000 Kragujevac, Serbia

Place of manufacture: PRIZMA d.o.o.
Zorana Đinđića 13, RS-34000 Kragujevac, Serbia

This certificate is granted subject to the SIQ's rules on product certification. SIQ certifies the conformity of the products with the requirements of the listed standards.

Ratings: 100-240 Vac; 50-60 Hz; 55 VA, Class I

Standard: EN 60601-1:2006 + A1:2013 + A12:2014
EN 60601-1-6:2010 + A1:2015
EN 62366:2008 + A1:2015


Test report: T223-0245/16, dated 2016-04-13
T223-0246/16, dated 2016-04-13
T223-0247/16, dated 2016-04-13

Remarks: This certificate shall apply to the products identical to the tested sample and shall remain valid for the period of 3 years until 2019-04-13 or until the validity date of the listed standards, whichever occurs earlier.

Date: 2016-04-13

Authorized signature:

Bojan Pečavar



Only integral publication of this certificate is allowed. This certificate may only be reproduced in its entirety and without any changes. On request SIQ will give information about the validity of the certificate.